Vespa – RFPulse
User Manual and Reference

Version 0.2.0

Release date: May 2nd, 2011

Developed by:

Brian J. Soher, Ph.D.
Philip Semanchuk

Duke University Medical Center,
Department of Radiology, Durham, NC

Karl Young, Ph.D.
David Todd, Ph.D.

University of California, San Francisco
Department of Radiology, San Francisco, CA

Developed with support from NIH, grant # EB008387-01A1

Table of Contents

Overview of the Vespa Package 4
Introduction to Vespa-RFPulse 5
Case Studies in RF Pulse Design 7
Using RFPulse – A User Manual 10
1. How to launch Vespa-RFPulse 10
2. Quick Guide – The Nuts and Bolts of RFPulse 12
3. The RFPulse Main Window 14
4. The Pulse Project Window 15
5. The Pulse Project Tab 16
5.1 Loading an existing pulse project 17
5.2 Running a new pulse project 17
5.4 Visualizing Pulse Project Results 18
6. Management Dialogs 20
6.1 Manage Pulse Project dialog 20
6.2 Manage Machine Settings Templates dialog 20
7. Results Output 22
7.1 Plot results to image file formats 22
7.3 Plot results to vector graphics formats 22
Appendix A. RFPulse Design 23
A.1 What is under the hood? 23
A.1.1 Vespa-RFPulse Basic Concepts 23
A.1.2 Pulse Projects 23
Appendix B. RFPulse Transforms 26
B.1 Basic Info Tab 26
B.1.1 Tab Diagram 26
B.1.2 General Usage 26
B.1.3 Widgets and Parameters 26
B.1.4 Algorithm Applied at Run Time 27
B.2 SLR Tab (create transformation) 28
B.2.1 Tab Diagram 28
B.2.2 General Usage 28
B.2.3 Widgets and Parameters 28
B.2.4 Algorithm Applied at Run Time 29
B.3 Hyperbolic-Secant Tab (create transformation) 30

B.3.1 Tab Diagram 30
B.3.2 General Usage 30
B.3.3 Widgets and Parameters 30
B.3.4 Algorithm Applied at Run Time 31
B.4 Interpolate-Rescale Tab (general transformation) 32
B.4.1 Tab Diagram 32
B.4.2 General Usage 32
B.4.3 Widgets and Parameters 32
B.4.4 Algorithm Applied at Run Time 32

Overview of the Vespa Package

The Vespa package enhances and extends three previously developed magnetic resonance spectroscopy (MRS) software tools by migrating them into an integrated, open source, open development platform. Vespa stands for Versatile Simulation, Pulses and Analysis. The original tools that have been migrated into this package include:

- GAVA/Gamma - software for spectral simulation
- MatPulse – software for RF pulse design
- IDL_Vespa – a package for spectral data processing and analysis

The new Vespa project addresses current software limitations, including: non-standard data access, closed source multiple language software that complicates algorithm extension and comparison, lack of integration between programs for sharing prior information, and incomplete or missing documentation and educational content.

Introduction to Vespa-RFPulse

Vespa-RFPulse is a graphical control and visualization program written in the Python programming language that provides a user friendly front end for calculation of Shinnar-LeRoux (SLR) frequency selective RF pulses, and many other calculations and manipulations. Although RFPulse is meant to be intuitive and is entirely menu-driven, a cursory reading of the information below is recommended. A description of the original MatPulse program (beta version 1.0) is provided in reference 1 (below). The RF pulses and gradient waveforms created by RFPulse follow the general procedures and nomenclature provided in references 2 and 3. Along with SLR based pulses, analytical pulse shapes, such as hyperbolic secant, are also available, as are additional transformation steps such as pulse integration and re-scaling.

1. Matson, G.B. An integrated program for amplitude-modulated RF pulse generation and re-mapping with shaped gradients. Magn. Reson. Imaging 12, 1205-1225, 1994.

2. J. Pauly, P. Le Roux, D. Nishimura, and A. Macovski, "Parameter Relations for the Shinnar-Le Roux Selective Excitation Pulse Design Algorithm", IEEE Trans Med Imaging 10: 53 - 65 (1991).

3. S. Conolly, D. Nishimura, A. Macovski, and G. Glover, "Variable-Rate Selective Excitation", J Magn Reson 78: 440 - 458 (1988).

What can RFPulse do?

1) Create new RF pulse projects from a list of design algorithms
2) Store pulse projects and their design parameters into a database
3) Re-load previous pulse projects
4) Display pulse results for each step of the design process in a flexible plotting tool
5) Compare side-by-side results from one or more pulse projects
6) Output results in text or graphical format, including MR manufacturer platform formats
7) Export/Import Vespa pulse projects to/from other users
8) Open source test bed for your own algorithms and transformations.

What is an RF pulse project? A ‘pulse project’ consists of one or more design steps. Each pulse project has a single creation step. You can further refine the pulse by adding transformation steps. Each step of the design process (creation plus other transformations) contains time and frequency domain results for the RF pulse up to that point. These results can be visualized in plots to the GUI at any time. Changes to any given step in the design process trigger a “downstream” calculation of all subsequent transformations.

You can open multiple pulse projects and compare them side-by-side. Any pulse project tab can be copied into a new pulse project tab. This allows you to make minor changes to a copy of an RF pulse in order to see the effects on the final results.

RFPulse can create (and expects to work with) four separate 'vectors' which you will see mentioned and explained in more detail in later sections:

The first, **B1**, represents an amplitude modulated RF pulse.
The second, **B2**, represents a re-mapped pulse (reshaped for use with shaped gradients).
The third, **G2**, represents the shaped gradient waveform for use with **B2**.

The fourth, **F2**, represents the frequency offsets for an offset slice (used with **B2** and **G2**). In addition, the constant gradient for a conventional selective pulse is labeled **G1**.

The following chapters run through the operation of the Vespa-RFPulse program both in general and screen by screen.

In this manual, command line instructions will appear in a fixed-width font on individual lines, for example:

```
~/Vespa-RFPulse/ % ls
```

Specific file and directory names will appear in a fixed-width font within the main text.

Online Resources:

The Vespa project and each of its applications have wikis with extensive information about how to use, and develop new functionality for, each application. These can be accessed through the main portal at

http://scion.duhs.duke.edu/vespa/

Case Studies in RF Pulse Design

Although your MRI instrument has a slew of RF pulses installed and available for use in new experiments, there are many reasons why you may want to consider designing new RF pulses. The figure below shows a generic pulse profile to define terms used to describe a pulse profile.



The following sections list some of the reasons why you may want to consider designing your own RF pulse designs.

Case 1 – Improved Selectivity

Many MRI pulses are designed with short TE sequences in mind, and the pulses are designed with a reduced length (at the expense of increased transition bands which creates some loss in selectivity and sensitivity) to accommodate short TEs. Somewhat longer pulses that provide for shorter transition bands may provide improved S/N for signals with longer T2s. In addition, pulse profiles with minimal transition bands are helpful for slice selection without slice gaps. RFPulse provides menus for designing longer RF pulses to reduce the transition bands.

Case 2 – Reduced Contamination

Especially for spectroscopy, contamination from signal from outside the excitation bandwidth can be a problem (e.g., leading to lipid or water signals from outside the voxel interfering with spectroscopic signals from within the voxel. Selective pulses designed for improved out-of-band (stopband) suppression can reduce this problem. The SLR pulse design tab in RFPulse enables the stopband suppression to be selected, and facilitates exploring the tradeoffs involved between increased stopband suppression, increased transition band, and increased pulse length.

Case 3 – Observation of Spectroscopic Signals Close to Water

Most MRI instruments use Gaussian pulses for water suppression. However, Gaussian-shaped pulses generate large transition zones, leading to suppression of other spectroscopy signals

close to water. The use of pulses with improved selectivity (e.g., SLR pulses as designed with RFPulse) can reduce this problem.

Case 4 – Lowered Peak Voltage Pulses

The peak voltage (or power) able to be applied by the MRI instrument to the coil is limited. This can prevent a desired RF pulse from being implemented on the instrument. For example, spin echo pulses require 3 to 4 times the peak voltage required by a 90 degree pulse producing the same bandwidth. In addition, hyperbolic secant inversion pulses require high peak voltage. There are several approaches to lowering the peak voltage of these pulses.

Spin echo pulses: An RF pulse can be expressed in terms of a pair of polynomials, designated A and B (see e.g. the paper by Pauli et al. cited above). For an SLR pulse, the B polynomial represents the magnitude of the selection profile that will result from the pulse, while the A polynomial contains the phase information for the pulse. For a so-called minimum phase pulse, the roots of the A polynomial lie within the unit circle in the complex plane.

Reflection of one of the roots of the A polynomial across the unit circle - altering the value of the root from z to 1/(z*) – doesn't alter the profile produced by the pulse, but may lower the peak voltage required. Typically, the peak voltage can be reduced to around 0.6 of the original voltage required. While the peak voltage is decreased, the price to be paid is that the SAR for the pulse is increased.

A second approach to lowering the peak voltage of a spin echo pulse is to use the VERSE technique (see e.g. the paper by Matson cited above), which can be thought of as lengthening the center portion of the pulse to enable the voltage of this section of the pulse to be reduced. The FOCI technique is similar, and can be considered as a particular implementation of the VERSE method. RFPulse incorporates a flexible version of the VERSE method, labeled remapping, which enables the RF pulse and associated gradient to be altered to lower the peak voltage of the pulse. While the RFPulse version does not necessarily increase the overall length of the pulse, depending on the parameters used, it may increase the SAR.

Hyperbolic secant inversion pulses: The hyperbolic secant pulse represents an adiabatic pulse, where the performance of the pulse does not change once the total power of the pulse is above a certain threshold. The advantage of this type of pulse is that it can perform well even in an inhomogeneous B1 field.

The voltage shape of the hyperbolic secant pulse can be expressed as $\mathrm{sech}(\beta\tau)$, where τ is the normalized time extending from -1 to 1. This shape produces a high voltage at the center of the pulse. Garwood et al. have shown that, by using a shape function of the form $\mathrm{sech}(\beta\tau^n)$, the peak voltage of the pulse can be reduced, while the pulse retains its adiabatic character.

The original hyperbolic secant pulse is thus designated HS1, while the reduced voltage pulses are designated HSn, where n indicates the power to which τ is raised. RFPulse enables HSn pulses to be designed over a range of values of n. Although larger values of n lower the peak voltage, the selectivity of the pulse is degraded at higher values of n.

Case 5 – Reduced SAR

Due to the high peak voltage required, both spin echo and hyperbolic secant pulses generate high SAR, and lower SAR versions of these pulses may be desired. For spin echo pulses, the RFPulse remapping (discussed in the previous section) can be employed with parameters chosen to lower the SAR (and lengthen the pulse duration) of a spin echo pulse. For a hyperbolic inversion pulse, the pulse can be implemented as an HSn pulse (discussed in the previous section) to lower the SAR.

Case 6 – More Exotic Pulses

Examples include multiband, increased bandwidth, concatenated, designed with immunity to B1 inhomogeneity, etc. Future additions to RFPulse will include a number of additional RF pulse design methods that enable a wide variety of additional designs to be implemented. These will include:

Multiband pulses: in which two separate selection bands are included within a single pulse design. For example, in a spectroscopy editing experiment, editing may be desired simultaneously in two separate regions, or suppression in two regions (e.g., water and lipid).

Increased bandwidth pulse designs: useful in spectroscopy to suppress the "two compartment artifact', which causes signal loss for J-coupled metabolites unless the experiment is conducted with very short TE.

Special concatenated RF pulses: used in a variety of experiments. For example, concatenated RF pulses are used in arterial spin labeling experiments such as the TILT experiment and in so-called pseudo continuous arterial spin labeling experiments.

B1 insensitive pulse design: Particularly at high magnetic field, the B1 field becomes inhomogeneous, and it may be desirable to use RF pulse designs that perform uniform tipping even in the presence of inhomogeneous RF fields. RFPulse provides optimization methods to obtain such designs.

Using RFPulse – A User Manual

This section assumes Vespa-RFPulse has been downloaded and installed. See the Vespa Installation guide on the Vespa main project wiki for details on how to install the software and package dependencies. http://scion.duhs.duke.edu/vespa.

In the following, screenshots are based on running Vespa-RFPulse on the Windows OS, but aside from starting the program, the basic commands are the same on all platforms.

1. How to launch Vespa-RFPulse

The usual case: If you installed RFPulse by downloading and unzipping the package and by running "python setup.py install", then you should already have an Icon on your desktop called RFPulse. Double clicking on this icon will launch the application.

Shown below is the Vespa-RFPulse main window as it appears on first opening. No actual RF pulse project windows are open, only the 'Welcome' banner is displayed.



Use the RFPulse menu to open an existing pulse project or to create a new project for designing an RF pulse.

Shown below is a screen shot of a Vespa-RFPulse session with two pulse project tabs opened side by side for comparison. The functionality of all tools, algorithms and transformation will be described further in the following sections.


RFPulse - Basic SLR 64pt
Pulse Project Transformations Management View Help
Pulse Project 1
Pulse Params
Tip Angle: 90.0 degrees
Time Steps: 64 increments
Duration: 8.0 milliseconds
Bandwidth: 1000.0 Hz
Separation: 0.0 Hz
Single or Dual Band
Single Band Dual Band
Non-Coalesced Phase Type
Linear
Filter
SLR (Remez) Least-Squares
Passband Ripple: 1.0 Percent
Reject Ripple: 1.0 Percent
Run
Plot Control
Freq Profile Abs Profile
Ext Profile Ext-Absolute Profile
Waveform
Usage Type: Excite
Basic Info SLR
Frequency Profile
Absolute Frequency Profile
RF Pulse Waveform
Mx/My [normal]
Pulse Project 2
Pulse Params
Tip Angle: 90.0 degrees
Time Steps: 64 increments
Duration: 5.0048 milliseconds
Bandwidth: 1000.0 Hz
Separation: 0.0 Hz
Single or Dual Band
Single Band Dual Band
Non-Coalesced Phase Type
Linear
Filter
SLR (Remez) Least-Squares
Passband Ripple: 1.0 Percent
Reject Ripple: 1.0 Percent
Run
Plot Control
Freq Profile Abs Profile
Ext Profile Ext-Absolute Profile
Waveform
Usage Type: Excite
Basic Info SLR
Frequency Profile
Absolute Frequency Profile
RF Pulse Waveform
Mx/My [normal]
X = -6.600
Y = 0.92185
Value = 0.00962

2. Quick Guide – The Nuts and Bolts of RFPulse

The creation and modification of RF pulses for use in MR experiments is as much an art as a science. We hope that the modularity and flexibility of design steps as presented in RFPulse will allow and encourage users to play with, and better understand, the effects of changing parameter values in each step of the design. Also, since each step is on a separate tab with in the pulse project, you can see the results of each step.

To facilitate your initial usage, we offer here a quick description of how to get started using and learning from RFPulse.

We start with the assumption that you have already installed and launched the RFPulse application. You should see the Welcome tab displayed in a Notebook with no pulse project tabs displayed.

The easiest interaction requires just a few basic steps.

- From the **PulseProject** menu, select New to create a new pulse project tab in the Notebook
- In the Basic Info tab, fill in the project name and investigator fields.
- Select **Transforms→Create→SLR** and a new "SLR" tab appears.
- Hit the **Run** button and a time domain RF pulse waveform is displayed in the right panel.
- (optional) select other plot options at the bottom of the tab to see other results plots.
- (optional) vary parameters and hit **Run** again to observe desired pulse performance.
- Select **PulseProject →Save** menu item to store your results to the Vespa database so you can re-open them later.

From here you can change parameters on the SLR creation tab to refine the shape, duration or other features of the pulse. You can also select other transformations from the Transformations menu. Each transformation appears on a new tab and allows you to further refine your pulse. (The motivations for manipulating RF pulses have been briefly discussed in the Case Studies section).

There are two types of transformations, "create" transformations and "general" transformations. As shown above, pulse projects start with a create transformation. Create transformations are the algorithms by which an initial RF pulse is designed (e.g. Shinnar-LeRoux, hyperbolic secant, etc.). Each pulse project has just one create transformation. The create transformations are selected from the **Transforms→Create** menu. General transformations are listed below the **Create** submenu. General transformations operate on the results of the previous transformation (i.e., the transformation in the tab to the left).

Notes

1. Another typical workflow might be to load a saved pulse project from the database, add, edit or delete general transformations, and then save the changed results back to the database.
2. A third typical workflow might be to load a saved pulse project from the database, and copy that into a new pulse project by using the **PulseProject→Copy Tab to New Tab** menu item. Then modify the RF pulse in the copied tab and compare the results plots to the original pulse project results.

3. Changes can be made to parameter settings in any transformation tab in a pulse project. When you click the Run button, RFPulse calculates the effects of these changes in all transform tabs “downstream” from the transformation tab changed. **Note** – as of this writing, there is no “undo” functionality in the transformation tabs in the application. Although, for existing pulse projects that have been loaded into a tab, changes to that project are not saved to the database until the user selects the **PulseProject →Save** menu manually.

3. The RFPulse Main Window

This is a view of the main Vespa-RFPulse user interface window. It is the first window that appears when you run the program. It contains a tabbed window into which pulse projects can be loaded, a menu bar and status bar. One or more pulse projects can be loaded at a time. Each project is displayed in its own tab located at the top of the window. Each project contains input data and results from one pulse project design. As described above, a pulse project is a group of transformation steps that create and modify an RF pulse. Each pulse project tab contains a group of tabs at the bottom of the page that represent the transformations that make up the project. Transformation tabs contain the input values and results for that step of the pulse design.



The pulse project window is initially populated with a welcome text window, but no pulse projects are loaded. From the RFPulse menu bar you can 1) load a previously run pulse project from the Vespa database into a tab, or 2) create a new pulse project and set it up and run it. In either case a tab will appear at the top of the window for each pulse project that is loaded or created. The Management menu allows users to access pop-up dialogs to create, edit, view, delete and import/export pulse projects and Machine Settings (which will be explained fully later).

The status bar provides information about where the cursor is located within the various plots and images in the interface throughout the program. It also reports short messages that reflect current processing while events are running.

On the Menu Bar

PulseProject→New	Opens a new pulse project tab.
PulseProject →Copy Tab to New	This will open a new pulse project tab and populate it with the same values that are listed in the current pulse project. This is a short cut for varying design parameters to get different results while retaining the ability to compare back to a previous results set without having to save them both to the data base.
PulseProject →Open	Runs the pulse project Browser dialog, from which you can choose a pulse project to open from the database to open.
PulseProject →Save	Saves the pulse project in the current tab to the database. **Note** - pulse project results are not automatically saved to data base after a Run button is hit.
PulseProject →Close	Closes current pulse project tab. Will prompt for save if necessary.
Transforms →Create →SLR	Appends an SLR (create transformation) tab to current pulse project
Transforms →Create →Hyperbolic-Secant	Appends an Hyperbolic-Secant (create transformation) tab to current pulse project

Transforms
→Interpolate, Filter, Rescale Appends an Interpolate, Filter, Rescale (general transformation) tab to current pulse project.

Management
→Manage Pulse Projects Launches the Manage pulse projects dialog. Allows user to view, clone, delete, import and export pulse projects.

Management
→Manage Machine Settings Templates Launches the Manage Machine Settings dialog. Allows user to create, edit, view, and delete templates for machine settings information.

View→<various> Changes plot options in the selected transformation tab of the active pulse project tab, including: display zero line, turn x-axis on/off or choose units, selecting data type and various output options for all plot windows.

Help→User Manual Launches the user manual (from vespa/docs) into a PDF file reader.

Help→RFPulse/Vespa Online Help Online wiki for the RFPulse application and Vespa project

Help→About Giving credit where credit is due.

4. The Pulse Project Window

The pulse project window is an “advanced user interface” widget (AUINotebook). What that means is a lot of flexibility in how the window can be used: Multiple tabs can be opened up inside the window. They can be moved around, arranged and “docked” as the user desires by left-click and dragging the desired tab to a new location inside the notebook boundaries. In this manner, the tabs can be positioned side-by-side, top-to-bottom or stacked (as show in Sections 1 and 4). They can also be arranged in any mixture of these positions.



The pulse project window can be populated with one or more pulse project tabs, each of which contains the results of one pulse project. Pulse project tabs can be closed using the X box on the tab or with a middle-click on the tab itself. When a pulse project Tab is closed, the pulse project is removed from memory, but can be reloaded from the database at a future time - assuming it was previously saved.

5. The Pulse Project Tab

A pulse project tab is one page in the pulse project window. Each tab contains one entire pulse project design. A pulse project tab can be used to run a new pulse project and view the results of that design. It can also be used to load an existing pulse project from the database to view results, or to change parameters or other modifications to the pulse project.

Each pulse project tab will contain sub-tabs at the bottom of the page that describe the transformations used to design the RF pulse. There will typically be two or more of these "transformation tabs". The first tab is standard to all pulse projects and is called "Basic Info". The second tab is always a "create" transformation tab such as "SLR" or "Hyperbolic-Secant" (see Appendix B). Additional tabs, containing "general" transformations such as Interpolate-Filter-Rescale (see Appendix B) can be appended after the "create" transformation tab. The pulse project tab can be resized by adjusting the RFPulse application window and each transformation tab has a vertical splitter bar through the middle that can be used to resize the right and left sides. Parameters for each transformation tab are displayed on the left side of the tab, the right side of the tab contains a plotting canvas to visually display the current RF pulse waveform and profile results. Typically, when a transformation tab is added to the pulse project, there are no results to be displayed until the **Run** button has been clicked.

A New pulse project is typically created, set up and run. After this, other changes to parameters can be made and the Project re-run until the user is satisfied with the pulse performance. Results from running a pulse project are only saved to the database when specifically requested by the user. Transform tabs are updated to display results after each click on the Run button. If there are transformation tabs "downstream" from the tab where the **Run** button was hit, the program will attempt to automatically propagate changes through those tabs as well (ie. essentially automatically triggering their Run functions). Pulse projects can be modified in any tab multiple times. Again results are only saved upon the user specifically calling **Pulse Project→Save**.

All saved pulse projects start life as "private". That means they're only in your database; no one else has seen/accessed them. Once you export a pulse project to share with another user, the pulse project is designated as "public". That means the pulse project result has been shared with the world. Public objects are "frozen" and become mostly unchangeable. Once a private pulse project has become public, it can never become private again.

There are two ways that a public pulse project can be re-used. First, cloning a public pulse project (using the Manage Pulse Projects dialog) will create a new, private pulse project in the database with exactly the same properties as the original (but a different UUID and name). This clone can then be loaded into the pulse project window and modified. Second, any pulse project can be loaded into the pulse project window, whether it is public or private. From there, the **Pulse Project→Copy to New Tab** menu item will create a new pulse project tab with a copy of the current tab. This copy will be private and ready to be modified. Note that an important difference between cloning and copying tabs is that the clone is automatically saved into the database on creation, while the copied tab is only created in memory until the user specifically saves it.

The View menu on the main menu bar can be used to modify the display of the plots in the active pulse project tab. The resulting modifications only affect the settings in the active pulse project tab. More specifically, only the plot options in the selected transformation tab in the active pulse project tab are affected. These options are preserved as you switch from transformation tab to transformation tab. The following lists the functions on the View menu item:

On the Menu Bar

View (this menu affects the plots in the currently active PulseProject/Transform tab)

→Show ZeroLine	toggle zero line off/on in waveform and profile plots
→Xaxis →Show	turns off/on vertical grid lines and values along the x-axis
→Xaxis →[mm]	changes the units for values along the x-axis
→Xaxis →[kHz]	changes the units for values along the x-axis
→Data Type	select Real or Real+Imaginary in waveform and frequency profile plots
→Output→<format>	writes the plot panel to file as either PNG, SVG, EPS or PDF format

5.1 Loading an existing pulse project

The pulse project Browser dialog is launched from **Pulse Project→Open** menu which is shown below. A list of pulse project names is shown on the left. When a pulse project listed in the browser is clicked on once, its comment and other information are displayed on the right.

When the Open button is clicked (or a pulse project's name is double-clicked on), the program loads the information for that pulse project from the database into a pulse project object in memory. This object then creates a pulse project tab and sets up the transformation tabs that describes its creation. Note. in the case of a large pulse project, this may take a short but noticeable amount of time to finish.



5.2 Running a new pulse project

When a user selects the **Pulse Project→New** menu option, a new pulse project tab is created in the pulse project window and the Basic Info transformation tab loaded. This tab requires the user to type in a project name before the project can be saved. A project investigator and/or a comment can optionally be added.

A list of available transformations is can be seen and selected from the **Transformations** menu. Create transformations are found under the **Tranformations→Create** sub-menu. All the other transformations directly under the **Transformations** menu are general transformations. A detailed description of each transformation and its general usage can be found in Appendix B.

As noted previously, a 'pulse project' object consists of one or more transformation objects. Each pulse project object must start with a single create transformation but can contain zero or more general transformation tabs. Each transformation object contains results for that step in the RF pulse design.

For more detail on how to get started on creating and manipulating RF pulses, see **Section 2 – Quick Guide and the Case Study** section on pulse design in this manual.

5.4 Visualizing Pulse Project Results

Transform tabs - displayed inside a pulse project tab - each contain a plot panel on the right hand side of the tabbed window. This plot panel 'canvas' can display any (and all) of the five results plots, in any combination, by selecting from the check boxes in the **Plot Control** panel on the left side of the tabbed window. Turning a check box on or off adds or removes the plot from the figure on the right.

In order to modify the data displayed, and other display parameters, use to the **View** menu item.

The plots displayed in each transformation tab are the cumulative results for all transformation tabs to the right of, and including, the current tab.

When a transformation tab is added to a pulse project, there may not be results plotted until the Run button is clicked. Changing transformation parameters on the left panel does not typically change the results plotted until after the Run button is hit.



On the Plot Control Panel

Freq Profile (checkbox) Selects RF pulse frequency profile to be plotted. This result comes from running the RF waveform through the Bloch equations to calculate the complex RF profile at ±1 times the Nyquist frequency and at the calc_resolution set in the Basic Info tab.

Abs Profile	(checkbox) Selects RF pulse absolute frequency profile to be plotted. The magnitude plot of the frequency profile plot described above.
Ext Profile	(checkbox) Selects the RF extended frequency profile to be plotted. This result comes from running the RF waveform through the Bloch equations to calculate the complex RF profile at ±4 times the Nyquist frequency and at the calc_resolution set in the Basic Info tab. This allows the user to check a wider bandwidth range, albeit at a coarser spectral resolution than for the non-extended profile. This allows the user to view out-of-ban excitation.
Ext-Absolute Profile	(checkbox) Selects the RF extended absolute frequency profile to be plotted. The magnitude plot of the extended frequency profile plot described above
Waveform	(checkbox) Selects the time domain RF pulse waveform to be plotted.

The **Usage Type** drop list widget requires a more careful explanation. For any given RF pulse, running the complex time waveform through the Bloch equations results in a description of the magnetization along the Z axis as well as in the X,Y plane. Depending on what the pulse will be used for, the user will want to check the pulse performance of either the M_z or $M_{x,y}$ and sometime both directions. Changing the Usage Type widget allows the user to specify what is plotted along the vertical axis (a.k.a. the y-axis) in each plot in the figure.

Usage Type effects on Y-axis Choice

Excite	The M_x (real part) and M_y (imaginary part, if turned on) magnetization is plotted in all profile plots, with the assumption of initial magnetization along the z axis.
Inversion	The M_z magnetization is plotted in the real part and zeros in the imaginary part (if turned on) in all profile plots, with the assumption of initial magnetization along the z axis.
Saturation	The $-M_z$ magnetization is plotted in the real part and zeros in the imaginary part (if turned on) in all profile plots, with the assumption of initial magnetization along the z axis.
Spin Echo	The M_x (real part) and $-M_y$ (imaginary part, if turned on) magnetization is plotted in all profile plots, with the assumption of initial magnetization along the y axis

The mouse can be used to zoom in on the X and Y dimensions in all plots and to set vertical reference spans along the x-axis. The left mouse button draws a box which designates the region into which to zoom the plot. Click and drag the left mouse button in the window and a yellow box will appear. Drag the mouse to size the zoom box appropriately; release the left button and the plot will zoom into that region. XRange, YRange, plot value, and ΔX and ΔY values will be shown in the status bar while dragging. Click and release the left mouse button in place and the plot will zoom out to its max setting.

In a similar fashion, two vertical cursors can be set inside each plot window. Click and drag the right mouse button then release to set the two cursors anywhere in the window. This Cursor Span will display as a light gray span. Click and release the right mouse button in place and the cursor span will be turned off.

6. Management Dialogs

The Management dialogs allows the user to create, delete, edit, import, export or view pulse projects and machine settings templates. These dialogs thus allow the user to manage the data in the RFPulse database. It also provides the means for users to share information between themselves via XML files created using the Import/Export functions. Note. Due to the highly system specific nature of MR scanner specifications, machine settings templates are not Import/Export compatible.

6.1 Manage Pulse Project dialog

Access this dialog by clicking on the **Management→Manage Pulse Projects** menu item. The dialog opens and blocks other activity until it is closed. An example of this dialog is shown in the figure. Pulse project names are listed in the window on the right. Users may View, Delete, Clone, Import or Export pulse projects. These functions are summarized below.



View: Creates a brief textual description of the pulse project that is displayed in a native text editor for the platform being used.

Clone: Creates an exact copy of the currently selected pulse project into a new pulse project and saves it into the database. Only the unique ID is changes. This is typically used on pulse projects that have been designated as “public” in order to have a modifiable copy with which to work.

Delete: Removes the pulse project from the database.

Import: Allows the user to select an XML file that contains a pulse project. If the UUID in the file is unique, it is added to the pulse project database.

Export: The user selects a pulse project from the list. A second dialog allows the user to browse for the output filename, select if output should be compressed and allows an additional export comment to be typed in. Note that the action of exporting a pulse project (or other objects) caused it to be marked as “frozen” in the database. This means that no changes can be made. This is for the sake of consistency as results are shared. However, a frozen pulse project can still be deleted from the database if needed. This file can be imported into another Vespa-RFPulse installation using the Import function. If additional changes are desired a new pulse project, cloned from the original “frozen” object, can be created then modified.

6.2 Manage Machine Settings Templates dialog

Access this dialog by clicking on the **Management→Manage Machine Settings Templates** menu item. Actions that can be taken on the Metabolite dialog include, New, Edit, View, Clone, Set Default and Delete. An example of this dialog is shown below. The "Default" column indicates which template is used as the default settings for the machine settings object in a newly created pulse project object. The Set Default button is used to designate the currently selected template as the default template. Only one template can be set as the default at one time.



New: A dialog will pop up that gives the user a blank form to fill out. Fill in the settings in the widget fields and hit the OK or Cancel button. See the sample in the figure below.



Edit: The highlighted metabolite is opened in the machine settings editor. All fields are editable. The name is editable because pulse projects do not save specific machine settings templates. These templates are just a convenience tool to set these values in the pulse project object.

View: Similar to Edit but no fields are editable.

Clone: Select a template in the list, hit clone and a copy of that template is made that is now fully editable. The new template has the name of the original metabolite followed by the date and the word "_clone".

Delete: Any template may be deleted. This is because pulse projects do not save specific machine settings templates. These templates are just a convenience tool to set these values in the pulse project object.

Set Default: Used to designate the currently selected template as the default template that sets the machine settings in a new pulse project object.

7. Results Output

7.1 Plot results to image file formats

Results plots in each transformation tab can all be saved to file in PNG (portable network graphic), PDF (portable document file) or EPS (encapsulated postscript) formats to save the results as an image. The Vespa-RFPulse **View** menu lists commands that only apply to the active transformation tab in the active pulse project tab. Select the **View→Output→** option and further select either the **Plot to PNG, Plot to PDF** or **Plot to EPS** menu item. The user will be prompted to pick an output filename to which will be appended the appropriate suffix.

7.3 Plot results to vector graphics formats

Results in each transformation tab can be saved to file in SVG (scalable vector graphics) or EPS (encapsulated postscript) formats to save the results as a vector graphics file that can be decomposed into various parts. This is particularly desirable when creating graphics in PowerPoint or other drawing programs. At the time of writing this, only the EPS files were readable into PowerPoint.

The Vespa-RFPulse **View** menu lists commands that only apply to the active transformation tab in the active pulse project tab. Select the **View→Output→** option and further select either the **Plot to SVG**, or **Plot to EPS** menu item. The user will be prompted to pick an output filename to which will be appended the appropriate suffix.

Appendix A. RFPulse Design

A.1 What is under the hood?

A.1.1 Vespa-RFPulse Basic Concepts

This is a combination of logical concepts and constraints that determine how RFPulse works. These rules are enforced through the application and, to some extent, the database.

The main objects in the system are pulse projects, transformations, machine settings and master parameters. There are two types of transformations, “create“ and “general“, which will be described in more detail later. Pulse projects are the primary objects; everything else is secondary. Here's how they're related --



- Each pulse project has one or more transformations. Transformations are the building blocks of a pulse project design. The pulse project object contains a list of all transformation objects that are part of the RF pulse design. The pulse project also contains the machine settings and master parameter settings used.
- Each transformation step of the design contains the results of the RF pulse at that step, as well as the parameters used to create that result. The pulse project can be saved at any time, but changes made in the GUI will not be copied into internal objects and thus saved until after the “Run” button is hit (for a given transformation).
- Each pulse project contains exactly one “create” transform. This transformation is the first one selected by the user, and is positioned as the second transformation tab in the GUI after the standard Basic Info tab.
- Each pulse project can add any number of general transformations after the create transformation.
- Each pulse project has only one set of machine settings.
- Each pulse project has only one set of master parameters.

We expect users to share data via RFPulse’s export and import functions. For this reason, the RFPulse pulse project object has a universally unique id (UUID) rather than just ordinary integer ids. Due to the site specific nature of machine settings, these objects are not independently available for sharing through the import/export mechanism. However, the machine settings used in a given pulse project are stored in that project when it is exported and are thus available for use when it is subsequently imported.

A.1.2 Pulse Projects

Pulse projects are the main focus of the RFPulse application. A Pulse Project’s *raison d'etre* is to help a user design an RF pulse through a series of transformation steps. These

transformations first create and subsequently modify an RF pulse to achieve the desired results. This set of transformations is the pulse project's *transformation chain*.

Currently, that chain is implemented as a list of transformation objects. The first transformation object is always the Basic Info that describes the pulse project meta-data and allows the user to select the Machine Settings and Master Parameters for the pulse design. The second transformation is always a create transformation that implements an RF pulse design algorithm, such as Shinnar-Le Roux or hyperbolic-secant, that creates an initial RF pulse waveform result. Subsequent steps always contain a general transformation which uses the waveform result of the previous step as an input to be modified by the current transformation step.

The transformation chain design was selected to allow more modularity in RF pulse design and increased flexibility in the display of intermediate results. Each transformation tab in the pulse project has a manageable number of parameter widgets in its GUI. These can be modified and applied to the local set of results without having to run all previous steps. Results are plotted in the transformation tab to allow the user to visually inspect the effects of their settings for each design step.

The figure below is a visual representation of two steps in a transformation chain. Each transformation contains the parameters used in its algorithm, a set of results from running its algorithm, and a reference to the results in the previous transformation step (used as an input).



Transformations can be added or removed from the chain, typically by closing the associated tab in the GUI, with a few caveats. New transformation steps can only be added to the end of the current chain. Any general transformation can be closed/removed, however this triggers an automatic recalculation of the steps "downstream" of that transformation. The create transformation can not be removed unless all general transformations have been previously closed/removed. This is due to the fact that the general transforms would have no results on which to act without an initial create transformation.

The take-home lesson from this section is that the Vespa-RFPulse application provides a modular design platform for optimizing RF pulses. Various steps in RP pulse creation and modification can be included in a given design, or not. In the appendix that follows, we will specify what each transformation step included does, are and how they are typically used.

Other useful information:

- Anytime a tab's GUI is out of sync with the parameter values used for the last Run of the transformation, the tab's name will contain an asterix ('*') next to it. Hitting Run will clear this.

Appendix B. RFPulse Transforms

This section provides some basic information about the transformations (both “create” and “general” types) available in the RFPulse application. A description of the user interface and the parameters that can be set there is given. A brief overview of the algorithm applied when the Run button is hit is also given.

B.1 Basic Info Tab

B.1.1 Tab Diagram



B.1.2 General Usage

This is a special transformation tab, neither “create” nor “general” type, which the user must always fill in so that RFPulse can correctly identify the pulse project to/from the database. The name field MUST be filled in before any other transformation tabs can be added to the pulse project.

B.1.3 Widgets and Parameters

Name – text, must be a text string that is unique to all other pulse project names in the database

Investigator – text, (optional)

UUID and Created – not editable, just given FYI

Machine Settings – label, listing the name of the currently selected machine settings.

Edit – button, opens an editable dialog that displays the current machine settings for modification.

Calculation Resolution – integer, the number of points used in the Bloch equation calculations to create pulse profiles. More points results in a more finely sampled bandwidth but at the expense of longer calculation times.

Bandwidth Type – droplist,

B.1.4 Algorithm Applied at Run Time

There is no Run button on this tab. To proceed to the next step the user selects a “create” transformation from the **Transform→Create** menu. The create transformation tab will check to see if a Name has been typed into the Basic Info tab before letting the user Run the create algorithm.

B.2 SLR Tab (create transformation)

B.2.1 Tab Diagram



B.2.2 General Usage

This is a “create” transformation tab. The SLR tab provides for creation of computer optimized SLR shaped (amplitude-modulated) RF pulses typically denoted as **B1**.

B.2.3 Widgets and Parameters

Tip Angle – float, effective tip angle of the pulse in degrees. As of this writing, this parameter is limited to shallow tip angles (1-40 degrees), 90 degree or 180 degree angles. Note that the pulse is shaped for the specified tip angle. To see the effect of the pulse at the wrong power (that is, producing the wrong tip), the pulse must be re-scaled and the profile calculated with the Bloch Equations.

Time Steps – integer, number of steps in the pulse. The number of pulse steps is usually selected in powers of two to speed the calculations. The profile shape is not improved by increasing the number of steps beyond around 128. However, the smoothness of the pulse (and minimization of out-of-band effects) is improved with more steps. For 90 degree pulses the minimum number of steps recommended is 64 , while for 180 degree pulses the recommended minimum is 128. Note that the calculations are much more rapid for a small number of steps (e.g., 64 steps), and a larger number of steps may be selected after the other pulse parameters have been decided upon.

Duration – float, in milliseconds, total time of the pulse. Values selected may not be achievable due to the minimum dwell times set in the machine settings, in this case an achievable value will be suggested.

Bandwidth – float, in Hz, width of the pass band in Hz as measured at full width half max amplitude value.

Separation – float, in Hz (optional) only activated if Dual Band is selected. This is the separation in Hz between the two passbands as measured at full width half max amplitude values.

Single or Dual Band – radio button, sets whether one or two passbands are in the design of the pulse. Single Band produces conventional slice-selective pulses. The Dual Band selection can be used to produce a dual band pulse such as a dual band saturation pulse.

Non-Coalesced Phase Type – dropbox, selections 'linear', 'max' or 'min'. Refocused (for a linear phase pulse) is generally preferred for an observation or spin echo pulse, while Max Phase or Min Phase are preferable for inversion or saturation pulses

Filter – radio button, selects whether the SLR algorithm uses a Remez or least-squares optimization algorithm.

Passband Ripple – float, in percent, of the RF pulse profile. See notes below in 'Reject Ripple"

Reject Ripple – float, in percent, of the RF pulse profile. These 'ripple' parameters are used to estimate a transition band, which is the parameter actually used in the calculations. Although the ripple estimates are quite accurate for 180, 90, and very shallow tip pulses, they may be in error for pulses calculated for intermediate angles

B.2.4 Algorithm Applied at Run Time

When you hit the **Run** button …

The following description is based on that in (Handbook of MRI Pulse Sequences, Bernstein MA, Kevin F. King KF, Xiaohong JZ. Academic Press, 2004). The SLR algorithm was developed to solve the difficult inverse problem of finding what RF pulse to apply, given a desired slice profile and the initial condition of the magnetization. For small flip angles, the shape of an excitation pulse can be approximated by an inverse Fourier transform of the slice profile but this approximation breaks down for pulses in the range of 30-90° or larger.

Iterative numerical optimization methods can be used for large tip angles but the SLR method allows for a direct, non-iterative solution in certain situations. Characteristics such as RF bandwidth, pulse duration, flip angle, percent ripple in the passband, and percent ripple in the stopband can be specified as well as acceptable tradeoffs between these parameters, and the algorithm returns the exact RF pulse through a straightforward computational process.

The SLR algorithm uses two key concepts: the SU(2) group theoretical representation of rotations and the hard pulse approximation. Two typical ways of describing rotations in three-dimensional space are: 1) via 3 x 3 orthogonal rotation matrices and 3 x I vectors, i.e. via the special orthogonal 3D group SO(3), 2) via 2 x 2 unitary matrices, and 2 x 1 complex vectors called spinors, i.e. via the special unitary group SU(2) (Pauly et al.1991). These two representations are completely equivalent ways of describing macroscopic rotations such as those experienced by the magnetization vector. But the SU(2) representation offers considerable mathematical simplification and as a result is used in the SLR algorithm.

The second important conceptual component of the SLR algorithm is the hard pulse approximation (Pauly et al. 1991), which is useful in particular in the case of non-adiabatic pulses. The hard pulse approximation states that any shaped, or soft, pulse $B_1(t)$ can be approximated by a series of short hard pulses followed by periods of free precession. The larger the number of hard pulses used, combined with the resultant decrease in the duration of the free precision periods, the more accurate the approximation.

By describing rotations in the SU(2) representation and using the hard pulse approximation, the effect of soft pulses on the magnetization vector can be mathematically described by two polynomials with complex coefficients. The process that transforms from the RF pulse to the two polynomials is called the forward SLR transform. It is important that the inverse SLR transform also can be calculated. The inverse transform yields the RF pulse, given the two complex polynomials corresponding to the desired magnetization. In digital signal processing (DSP), these polynomials are filters for which there are well-established and powerful design tools available. In the SLR algorithm, the inverse SLR transform is used

in conjunction with finite impulse response (FIR) filter design tools to design RF pulses directly (Pauly et al. 1991).

Vespa-RFPulse uses these DSP tools to turn user input parameters into the appropriate complex polynomials and apply the inverse SLR transformation to generate a corresponding RF pulse.

B.3 Hyperbolic-Secant Tab (create transformation)

B.3.1 Tab Diagram



B.3.2 General Usage

This is a “create” transformation tab.

B.3.3 Widgets and Parameters

Total Rotation – float, in degrees,

Time Steps – integer, number of steps in the RF pulse

Dwell Time or Bandwidth – float, microseconds or Hz, respectively. Either the bandwidth or the dwell time must be provided (when one parameter is given, the other is automatically calculated). However, the width or dwell time calculations are only approximate

Cycles – float, number of cycles before truncating pulse.

Power(n) – integer, power of hyperbolic secant function

Sharpness(mu) – float, parameter that defines sharpness of function

Filter Type – droplist, apodization filter type – Hamming or Cosine

Filter Application – float, percentage, how much to apply the filter (0.0 to 100.0)

B.3.4 Algorithm Applied at Run Time

When you hit the **Run** button …

Hyperbolic secant pulses are adiabatic pulses that are typically used for inversion and are relatively insensitive to inhomogeneities in the B1 field and transmitter power setting, and result in very uniform frequency profiles. For adiabatic pulses both the amplitude and frequency of the pulse are typically varied. The pulses are adiabatic in that the amplitude is large enough and the frequency variation is slow enough (the adiabatic condition) that the longitudinal magnetization follows the direction of the effective magnetic field generated by the combined B0 and B1 fields. More specifically the adiabatic condition is that the precession of the magnetization vector about the effective field vector is more rapid than the change in angle of the effective field vector.

For a hyperbolic secant inversion pulse the initial effective field is aligned with B0 and the final effective field is aligned in the opposite direction. Hyperbolic secant pulses have the remarkable property that once B1 is strong enough, inversion is achieved over a frequency selective range (slice) such that further increases in B1 do not result in increases in flip angle.

Due to the frequency dependent phase that is generated by a hyperbolic secant pulse, they can not be used as refocusing pulses. A disadvantage of hyperbolic secant pulses and adiabatic pulses generally is that they typically involve high SAR, though as described in the case studies section above there are techniques for lowering the SAR associated for a given hyperbolic secant pulse.

Hyperbolic secant pulses constitute one of the few analytic solutions of the Bloch equations (Silver MS, Joseph RI, Hoult DI. Selective spin inversion in nuclear magnetic resonance and coherent optics through an exact solution of the Bloch-Riccati equation. Phys Rev A. 1985 Apr;31(4):2753-2755.) and are actually soliton or nondispersive solutions. They have the form:

$$B_1(t) = A(t)e^{-i\omega_1(t)t}$$
$$A(t) = A_0\,\mathrm{sech}(\beta t)$$
$$\omega(t) = -\mu\beta\tanh(\beta t)$$

where *A(t)* is the modulated, time dependent amplitude and $\omega(t)$ is the modulated, time dependent frequency, and μ and β are parameters.

RFPulse takes the specified user input parameters and generates a hyperbolic secant pulse of the above form.

B.4 Interpolate-Rescale Tab (general transformation)

B.4.1 Tab Diagram



B.4.2 General Usage

This is a “general” transformation tab.

B.4.3 Widgets and Parameters

Interpolate – checkbox, turns Interpolation algorithm on/off.

Interpol. Factor – integer, the interpolation requires that the dwell time be reduced by the interpolation factor. If the selected interpolation factor violates the min dwell time (set in the Machine Settings), an error is reported.

Rescaling – checkbox, turns Rescaling algorithm on/off.

Angle – float,

B.4.4 Algorithm Applied at Run Time

When you hit the **Run** button …